

## **93-7585 Meeting 10/15/98**

### **Attendees**

**Jim Stafford  
Stacy Wade  
John Skuba  
Luke Hinkle  
Bob Maxwell  
Jim Poulin  
Jack Gillespie**

**Discussed current status of special at Applied.**

**Will provide Jim Poulin with costs and drawings of the Pirani sensor so they can get a better handle on the project.**

**Calibration of the switch needs to be done after welding; however, it is not a difficult procedure and can be done at the service centers. Simply need system with closed loop pressure control.**

**Did not get a good feeling anyone was ready to pick up and run with this ball.**

**jws**